



## DECLARATION OF CONFORMITY

According to Annex III of the Directive 98/79/EC for in vitro diagnostic medical devices, of the European Council, we

- ◆ **Manufacturer**                    **Camp Medica SRL**  
Str. Vultureni nr. 80, Sector 4, Bucharest – Romania

declare under our sole responsibility that the following IVD medical device,

- ◆ **Product name**                    **See Now** rapid tests – see Annex 1,

meet the provisions of the European Directive 98/79/EC, concerning in vitro diagnostic medical devices, which apply to them.

- ◆ **Requests**                            European Directive no. 98/79/CE from October 27th, of the European Parliament Council for in vitro diagnostic medical devices
- ◆ **Conformity assessment route**                    Annex III (from European Directive 98/79/CE), sections 2 to 5
- ◆ **Classification**                    medical devices for in vitro diagnostic (IVDD)  
EDMA Code : 12 70 09

The following standards were used to prove the products conformity with the essential requirements of the above directive: **SR EN ISO 13485:2004.**

- ◆ **Standards Applied**                    EN 375: 2001, EN 980: 2003, EN 13612: 2002, EN 13640:2002, EN 13641:2002, SR EN ISO 14971:2007, SR EN ISO 13485:2004
- ◆ **Place, Date of Issue**                    Bucharest – Romania, 05.01.2011

**Quality Manager**  
Dr. Eng. Carolina Constantin

**Annex 1 – See Now rapid tests for toxicology**

CODE	Test name
SN 7.1	<b>See Now</b> Amphetamine (AMP)
SN 7.2	<b>See Now</b> Barbiturate (BAR)
SN 7.3	<b>See Now</b> Benzodiazepine (BZO)
SN 7.4	<b>See Now</b> Cocaine (COC)
SN 7.5	<b>See Now</b> Methylenedioxymeth-amphetamine (MDMA)
SN 7.6	<b>See Now</b> Methamphetamine (MET)
SN 7.7	<b>See Now</b> Morphine (MOP)
SN 7.8	<b>See Now</b> Methadone (MTD)
SN 7.9	<b>See Now</b> Opiate (OPI)
SN 7.10	<b>See Now</b> Tricyclic Antidepressants (TCA)
SN 7.11	<b>See Now</b> Marijuana (THC)
SN 7.12	<b>See Now</b> Phencyclidine (PCP)